

VELOCITY LATEX EXAMINATION



STANDARDS

Medical Device Directive	N/A	MDD 93/42/EEC
Personal Protective Equipment	✓	PPE 89/686 EEC
	✓	EC#1935/2004
Food Contact	N/A	ISO 22000 Food Safety Management
EN 455	1-4	EU Standard
EN 420	✓	EU Standard
EN 374	1-5	EU Standard
ASTM F1671	✓	US Standard
ASTM D6319/D3578	✓	US Standard
FSC™ Option	N/A	

SPECIFICATIONS

Weight (M Size)	5.5 grams	
Material	Latex	
Color	Natural White	
Inner Treatment	Polymer Coated	
Outer Surface	Finger & Palm Textured	
Length	Minimum 240 mm	
Sterile	Non-Sterile	
Thickness (Single Wall) mm	Middle Finger Min 0.13 ± .02 Palm Min 0.10 ± .02 Cuff Min 0.08 ± .02	
AQL (up to)	2.5 - 4.0	
Protein Content	≤200 µg/g	
Water Leak Test	Inspection Level GI AQL 2.5 - 4.0	
Visual & Aesthetics	Inspection Level S4 AQL Major 4.0 AQL Minor 6.5	
Physical Dimension	Inspection Level S2 AQL 4.0	
Physical Properties	Inspection Level S2 AQL 4.0	
Particulate Residue	≤2mg/glove	
Tensile Strength	Minimum 14 MPa Minimum 14 MPa	Before Aging After Aging
Elongation	Minimum 650% Minimum 500%	Before Aging After Aging
Force at Break	Median 6N Median 6N	Before Aging After Aging

